

STS Smart DP6	DP6 smart Probe for Luc	DP6 smart Probe for Ludium 3000		
	The STS simulated probe contains simulant placed on surfaces and	The STS DP6 smart probe houses the STS electronics, battery pack and detection system ready for connection directly to a real unmodified Ludlum 3000. The STS simulated probe contains a gas detection head which detects the presence of the simulant placed on surfaces and clothing, the resultant reading is displayed as counts per minute on the instrument Display.		
Dimensions (mm)	H 220	W 60	D 60	
Weight (KG)	0.9 KG			
Construction	Aluminium casing			
Display Type	N/A			
Backlight	N/A			
Battery	Powered from internal Li-ion battery. 12 hours approx. run time	Instrument turns on when Ludlum 3000 is switched on. Charge time approx. 6hours from supplied charger		
Detector	STS gas detectors situated behind perforated face plate			
Audio Output	Selectable on Instrument			
Alarm Thresholds	Selectable on Instrument			
Retained Functionality	All original instrument controls and switches retained	Software unchanged from real instrument.		
Connector	Standard BNC			
Operating & Storage Temperature	Operating temp 0 to +30C	Above 30C the stimulant will rapidly evaporate	Storage temp 0C to +40C	
Warm up time	30 seconds from switch on to ready.			
Available Probes	N/A			
Available Simulants	LS1 –liquid stimulant spray		se refer to MSDS sheets urther information	
Additional Information	used in hazardous environments sensitive electronics which may a stored in a clean and dry enviror environmental conditions such a conditioning units and outside ex is identifiable by a trainee.	The STS Smart DP6 is not designed to be intrinsically safe and therefore should not be used in hazardous environments. The units are not waterproof and contain delicate and sensitive electronics which may be caused to fail if exposed to moisture. Units should be stored in a clean and dry environment. Instrument response may be affected by environmental conditions such as excessive heat and humidity and by air flow, strong air conditioning units and outside exercises may need to be considered to ensure the stimulant is identifiable by a trainee.		

Safe Training Systems ltd Tel: +44 (0)1189 799591 Email: sales@safetrainingsystems.com

Web: safetrainingsystems.com Registered in England No.2654899 VAT no. GB572853808