STS Safe-Series Field Survey Instruments



Safe Training Systems Ltd.

| Instrument Name | STS Safe-FH40G | Telepole Probe | | |
|---------------------------------|---|---|---|--|
| | for use with the Thermo Telepole The Probe operates using an STS presence of a simulated radiation resultant reading displayed on the may be used in conjunction with | e STS Safe-FH40G Telepole Probe simulator is a simulated radiation probe designed | | |
| Dimensions (mm) | 170H | 48Dia | | |
| Weight (KG) | 0.25KG | | | |
| Construction | Moulded Plastic Case | | | |
| Controls | From FH40G | | | |
| Connection | Standard Thermo Telepole Connector | Probe is turned on- on connection to the FH40 | FH40 displays EXTERN when connected | |
| Battery | Powered from FH40G | | | |
| Detector | STS radio frequency Detector | | | |
| Operating & Storage Temperature | Operating temp 0 to +30C | Storage temp 0C to +40C | | |
| Warm up time | 10 seconds from switch on to ready. | | | |
| Available Sources | Safe-MiniSource, Safe-MiniSource Variable | Available in a range of activity levels | | |
| Additional Information | The STS Safe-FH40G is not designed to be intrinsically safe and therefore should not be used in hazardous environments. The units are not waterproof and contain delicate and sensitive electronics which may be caused to fail if exposed to moisture. Units should be stored in a clean and dry environment, batteries should be removed if storing for more than 4 weeks. Instrument response will be affected by environmental conditions such as the presence of large reflective surfaces, substantial metal structures and variable wall thicknesses. | | | |

Safe Training Systems ltd Tel: +44 (0)1344 483563 Fax: +44 (0)1344 485175 Email: sales@safetrainingsystems.com

Web: safetrainingsystems.com Registered in England No.2654899 VAT no. GB572853808